

**1. GAS CHROMATOGRAPH/AUTO SAMPLER/MASS SPECTROMETER & DATA SYSTEM  
FORENSIC SCIENCE CENTER  
FUNDING – BYRNE GRANT  
(Request for Bid sent to four vendors)**

RFB #04-19190

	Agilent Technologies	Agilent Technologies Alternate	Shimadzu Scientific	Quantum Analytics
1 ea. MSD Performance Turbo System	\$ 63,001.00	\$ 53,550.85	\$ 40,236.20	<b>\$ 53,551.00</b>
1 ea. Ion Gauge Controller for MSD	\$ 1,399.00	\$ 1,189.15	Included	<b>\$ 1,189.00</b>
1 ea. 44K, Ion Gauge installation	\$ 327.00	\$ 277.95	Included	<b>\$ 278.00</b>
1 ea. Network GC	\$ 15,357.00	\$ 13,053.45	Included	<b>\$ 13,053.00</b>
1 ea. AutoSampler Injector Tower	\$ 4,921.00	\$ 4,182.85	\$ 3,772.00	<b>\$ 4,183.00</b>
1 ea. Tower installation	\$ 199.00	\$ 169.15	Included	<b>\$ 169.00</b>
1 ea. AutoSampler 100 position tray	\$ 5,597.00	\$ 4,757.45	\$ 2,355.20	<b>\$ 4,757.00</b>
1 ea. 44K Tray installation	\$ 195.00	\$ 165.75	Included	<b>\$ 166.00</b>
1 ea. NIST MS Library	\$ 3,252.00	\$ 2,764.20	\$ 2,024.00	<b>\$ 2,764.00</b>
1 ea. 44K NIST Library installation	\$ 327.00	\$ 277.95	Included	<b>\$ 278.00</b>
Total	\$ 94,575.00	\$ 80,388.75	\$ 48,387.40	<b>\$ 80,388.00</b>
Brand	Agilent	Agilent	Shimadzu Scientific	<b>Agilent</b>
Delivery Date	4 wks ARO		30-45 Days ARO	<b>1 wk ARO</b>

On the recommendation of Jerry Phipps, on behalf of the Forensic Science Center, David Spears moved to **accept the low bid meeting specifications from Quantum Analytics for \$80,388.00**. Don Brace seconded the motion. The motion passed unanimously.

Note: The low bid from Shimadzu Scientific did not meet specifications because they did not have the component exchange ability with the current system now in use, as was required in the Request for Bid.