

1. GAS CHROMATOGRAPH/MASS SPECTROMETER – FORENSIC SCIENCE CENTER

FUNDING – FORENSIC SCIENCE CENTER

(Request for Bid sent to four vendors)

RFB #05-25400

	Varian	Agilent Technologies
1 ea. Flat Panel 19” Monitor	\$563.00	No Bid
1 ea. HP LaserJet 4100 Printer, including parallel printer cables	\$1,144.00	
1 ea. NIST02 MS Library & Program	\$2,805.00	
1 ea. CP 3800 Installation Kit	\$165.00	
1 ea. CP 8400 AutoSampler	\$8,108.00	
1200 Triple Quad GC MS System	\$110,044.00	
Programmable Head Pressure Control	\$2,804.00	
Grand Total	\$124,489.00	
Brand	Varian	
Delivery Date	6 weeks	

On the recommendation of Jerry Phipps, on behalf of the Forensic Science Center, Don Brace moved to **accept the bid from Varian for \$124,489.00**. Jennifer Dombaugh seconded the motion. The motion passed unanimously.

Note: This Request for Bid specified a triple quad system rather than a single quad system, which has been used in the past. The triple quad system will allow for a greater level of sensitivity in analysis. Varian is currently the only vendor who is manufacturing a triple quad system.